



82007-15

GOVERNMENT OF KARNATAKA
 KARNATAKA STATE DRUG LOGISTICS & WAREHOUSING SOCIETY**
 No. 1, Dr. Subbalakshmi Road, 4th Floor, Anand Road, Bangalore - 560079
 Phone: +91 98 2086221 Fax: +91 98 2086477

No. KSL/Equip/R K M C/Ventilator/130E/2019-2020

dated 04/04/2020


To,
 M/S. R K Medical Solution Company,
 Plot 84 main matthur Road,
 Near prakash Petrol Pump,
 Badrapur,
 New Delhi 110044

Sub- Deviations found in the delivered Ventilators.
 Ref- P.O No. KSL/Equip/R K M C/Ventilator/130E/2019-20
 dated: 24.03.2020 for Covid-19 pandemic.

We have received 15 Nos of Drager Ventilators on 14/04/2020. We have been trying to get your Firm to send a technical person for Pre dispatch inspection within there was no response from your end, a in-house inspection was done and the observations made by Health Equipment Officer KSLW5, Bangalore is as under.

1. No certificates (ISO, CE, FDA) /Brochures, Model with manufacturing details /Proof not available.
2. Ventilators found used and tampered, with broken / old stands.
 Ex: 1 Ventilator bearing SL No. ARNK-0251, EVITA-4, Version 04.22,
 11.1.2007 has the machine running hours of 46583 hours. Our F.O was clear that it should be a new machine.
3. As per the work order the Technical Specifications do not comply fully compared to the data sheet.
 1. Many essential modes /options not available: Cannot ventilate at Neonatal mode/Pediatric Mode- which is an essential mode for different Covid-19 patients as per the anaesthesia specialist opinion.
4. Suitable Technical Demonstration not provided from the OEM (Original Equipment manufacturer /Supplier).

The ventilator does not comply with the supply order, you are directed to take back the supplied ventilators at your own cost and your order stands cancelled.


 28/4/20
 Additional Director,
 KSLW5 Bangalore